

Initial Situation

3 sets of Cardio First Angel 2.0 devices were sent to Innosonian Europe for examination

Test Devices

Inotech Kunststofftechnik GmbH: Cardio First Angel 2.0
Brayden: Brayden Pro Manikin (IM16-R)

Test Person

Dr. Jonathan Smart, Global Product Development Director, Innosonian Europe

Test Date

25.11.2021

Review

Dr. Smart received 3 sets of Cardio First Angel 2.0 devices. He tested each device on a Brayden Pro (IM16-R) manikin connected to Brayden Online App (iOS iPad).

He tested each of the devices in turn (doing 60 chest compressions each time) and below.

Dr. Smart followed the instructions provided in the protective box.

During the compressions he listened for the two clicks from the device without looking at the real time feedback from the Brayden Pro Manikin and App – meaning, just guided by the 2 clicks of the Cardio First Angel 2.0 device.

Even without following the rhythm of the clicks (faster and slower) as recommended the depth-pressure relationship of the Cardio First Angel 2.0 devices have been achieved and are according to ERC requirements. Therefore, Innosonian states that their Brayden Pro Manikin (IM16-R) needs about 40 kg of elastic force to achieve 5 - 6 cm (from the Tomlinson et al paper / but also ERC Guidelines 2021)

Summary


All tested Cardio First Angel 2.0 devices reached out the recommended depth of chest deformation which is required in ERC guidelines 2021.

Attachments

3 Screenshots Measurement recording of the Brayden Pro Online App from the iPad- showing a summary spider chart, mean data and percentages for each device.

Overall

Score **100**



Detailed results

Summary	By CPR metric criteria	By Cycle			
Overall Performance 	Duration	0:00:30	Cycle	CCF	100 %
	Compressions		Count	Avg. Time-off	0 sec
	Avg. Depth	5.4 cm	60		
	Avg. Rate	109 / min	Hand position accuracy		100 %
Ventilations	Avg. Volume	-- ml	Total	Avg. Speed	-- times -- sec

• Chest compression performance is very good

More information (quality of CPR) is available via 'Brayden Online Expert'


Detailed results

Summary	By CPR metric criteria	By Cycle
---------	------------------------	----------

1 Compression depth	Too Shallow 0 %	Good 100 %	Too deep 0 %
2 Compression release	Incomplete release 2 %	Good 98 %	
3 Compression rate	Too slow 0 %	Good 100 %	Too fast 0 %
4 Score of CCF	100		
5 No. of compressions	Too few %	Good %	Too many %
6 Hand position	Incorrect 0 %	Good 100 %	
7 Ventilation volume	Too Little %	Good %	Too Much %
8 Ventilation frequency	Too slow %	Good %	Too fast %

Overall

Score **100**



Detailed results

Summary	By CPR metric criteria	By Cycle			
Overall Performance 	Duration	0:00:29	Cycle	CCF	100 %
	Compressions		Count	Avg. Time-off	
	Avg. Depth	5.5 cm	60	0 sec	
	Avg. Rate	112 / min	Hand position accuracy		100 %
Ventilations	Avg. Volume	-- ml	Total	Avg. Speed	-- sec

• Chest compression performance is very good

More information (quality of CPR) is available via 'Brayden Online Expert'


Detailed results

Summary	By CPR metric criteria	By Cycle
---------	------------------------	----------

1	Compression depth	Too Shallow 2 %	Good 98 %	Too deep 0 %
2	Compression release	Incomplete release 2 %	Good 98 %	
3	Compression rate	Too slow 0 %	Good 100 %	Too fast 0 %
4	Score of CCF	100		
5	No. of compressions	Too few %	Good %	Too many %
6	Hand position	Incorrect 0 %	Good 100 %	
7	Ventilation volume	Too Little %	Good %	Too Much %
8	Ventilation frequency	Too slow %	Good %	Too fast %

Overall

Score **100**



• Chest compression performance is very good

Detailed results

Summary	By CPR metric criteria	By Cycle			
Overall Performance 	Duration	0:00:28	Cycle	CCF	100 %
	Compressions		Count	Avg. Time-off	
	Avg. Depth	5.4 cm	60	0 sec	
Avg. Rate	113 / min	Hand position accuracy		100 %	
Ventilations	Avg. Volume	Total	Avg. Speed		
	-- ml	-- times	-- sec		

More information (quality of CPR) is available via 'Brayden Online Expert'

Detailed results

Summary	By CPR metric criteria	By Cycle
---------	------------------------	----------

1	Compression depth	Too Shallow 2 %	Good 98 %	Too deep 0 %
2	Compression release	Incomplete release 0 %	Good 100 %	
3	Compression rate	Too slow 0 %	Good 98 %	Too fast 2 %
4	Score of CCF	100		
5	No. of compressions	Too few %	Good %	Too many %
6	Hand position	Incorrect 0 %	Good 100 %	
7	Ventilation volume	Too Little %	Good %	Too Much %
8	Ventilation frequency	Too slow %	Good %	Too fast %